LEX25 Ex ia/ib certified thermocouple for zone 0, 1 and 2



Construction

ATEX certificate DEKRA 17ATEX0123 X

IECEx execution Ex ia IIC T6...T1 Ga

Ex ia/ib IIC T6...T1 Ga/Gb

Ex ib IIC T6...T1 GB

IECEx certificate IECEx DEK 17.0046X

ATEX execution II 1 G Ex ia IIC T6...T1 Ga

II 1/2 G Ex ia/ib IIC T6...T1 Ga/Gb

II 2 G Ex ib IIC T6...T1 Gb

Suitable for zone 0, 1 and 2 Ingress protection IP65 / IEC529

Ambient temp. range -45°C...+80°C, this can be limited depending on the materials applied

such as cables, connection heads, junction boxes and transmitters.

Sheath

Construction Mineral insulated (MgO)

MaterialSS316 standard, other materials possibleDiameterCustomer to specify (0.25-12.7mm)

Length Customer to specify

Hot junction

Isolated

Sealpot

Length 45mm

Diameter To suit cable diameter

Thermocouple

 $\label{eq:KTJ} \textbf{Calibration} \hspace{1cm} \textbf{K},\,\textbf{T},\,\textbf{J},\,\textbf{E},\,\textbf{N},\,\textbf{R},\,\textbf{S},\,\textbf{B},\,\textbf{U} \text{ or } \textbf{L}$

For duplex KK, TT, JJ, EE, NN, RR, SS, BB, UU or LL

Accuracy According to 60584-1

Lead wire

Type of cable PVC standard

Option PUR, Silicon or PTFE insulated cable

Cable length Customer to specify

Datasheet notification


This datasheet is purely indicative, build-up of model code may vary from

this datasheet depending on chosen model or execution!

45

Ordering code

*This datasheet is purely indicative, build-up of model code may vary from this datasheet.



H3-1 Rev.0 7/07





Thermo Electra B.V. Ph.
Weteringweg 10 Fa:
2641KM Pijnacker Err
The Netherlands http

Phone +31 (0)15 - 3621200 Fax +31 (0)15 - 3694082 Email sales@thermo.nl http://www.thermo-electra.com

