
LEX30 Ex d certified thermocouple for zone 1 and 2




Construction	
ATEX execution	II 2G Ex db IIC T6T1 Gb
ATEX certificate	DEKRA 17ATEX0123 X
IECEx execution	Ex db IIC T6T1 Gb
IECEx certificate	IECEx DEK 17.0046X
Springaction	>10mm standard, >40mm optional
Suitable for zone	1 and 2
Ingress protection	IP66 / IEC529
Ambient temp. limits	-45°C+80°C, this can be limited depending on the materials applied such as cables, connection heads, junction boxes and transmitters.
Connection head	
Material	Aluminium or Stainless steel 316
Cable entry	M20x1.5 standard, other threads possible. Two entries optional
Cover	Threaded with retaining chain
Fire Lock	At process side
Measuring insert	
Construction	Replaceable and springloaded
Diameter	6.0mm standard, other diameters possible
Туре	Mineral insulated (MgO)
Thermocouple	
Calibration	K, T, J, E, R, S, B, U, N or L
For duplex	KK, TT, JJ, EE, RR, SS, BB, UU, NN or LL
Accuracy	According to IEC 60584-1
Extension	
Туре	Nipple - union - nipple standard, other constructions possible
Material	Stainless Steel
Style / Connection	1/2"NPT standard, other threads possible
Length	150mm standard, other lengths possible
Datasheet notification	
	This datasheet is purely indicative, build-up of model code may vary from this datasheet depending on chosen model or execution!



Ordering code

*This datasheet is purely indicative, build-up of model code may vary from this datasheet.







Thermo Electra B.V. Weteringweg 10 2641KM Pijnacker The Netherlands Phone +31 (0)15 - 3621200 Fax +31 (0)15 - 3694082 Email sales@thermo.nl http://www.thermo-electra.com H3-1 Rev.0 7/07

